








# Clorox Healthcare® Equipment Compatibility Guide

		Bleach					Hydrogen Peroxide	Quat
<b>Vital Sign Monitors/ Consoles</b>								
		Clorox Healthcare® Bleach Germicidal Wipes	Clorox Healthcare® Bleach Germicidal Cleaners	Clorox Healthcare® Fuzion Cleaner Disinfectant	Dispatch® Hospital Cleaner Disinfectant Towels with Bleach	Clorox® Germicidal Bleach	Clorox Healthcare® Hydrogen Peroxide Cleaner Disinfectant Wipes	Clorox Healthcare® VersaSure Alcohol-Free Disinfectant Wipes
Manufacturer	Product							
Barco	Eonis® Diagnostic Displays							
GE Healthcare	Transport Pro							
	ApexPro Telemetry System					1:10-1:100		
Midmark	ApexPro Antenna Infrastructure, Transmitter and Receiver					1:10-1:100		
	IQvitals							
Phillips	IQvitals® Blood Pressure Cuffs							
	IntelliVue Patient Mx400/450/500/600/700/800							
	IntelliVue Telemetry Transmitters							
	IntelliVue Monitors							
Welch Allyn	Avalon Fetal Monitors							
	Connex Vital Signs							
	Connex Integrated Wall System					1:10		
	Accessory Cable Management Mobile Stand 4800/4900 Series							
	CareTemp Touch Free Thermometer					1:10		
	Propaq Encore® Blood Pressure Cuffs							
	Harvey™ Elite® Stethoscopes							
	LumiView™ Binocular Microscope							
MindRay	Digital MacroView™ Otoscopes and Halogen HPX® Illuminators							
	Accutorr 7 Vital Signs Monitor							
	Passport 12/Passport 8 Patient Monitor							
Mortara Instrument, Inc.	ELI 280/150c/250c 12-lead Resting Electrocardiograph					1:500 to 1:10		
Capsule Tech	SmartLinx Vitals Plus					10%		
Exergen	TAT5000 Temporal Thermometer							

## Legend

Product is mentioned by name in Cleaning & Care guide and is approved for use; if relevant, dilution is stated

Active ingredient/level is listed in Cleaning & Care guide as approved for use; if relevant, dilution is stated